



## INSTRUCTION FOR USE FOR SINGLE-USE MANOMETRIC WATER-PERFUSED AND/OR BAROSTAT CATHETERS

- This catheter is designed for **Single-Use Only**. Attempts to reprocess, resterilize, and/or re-use may lead to failure of the device and/or cross-contamination.
- Inspect catheter for any evidence of mechanical damage or imperfections. **DO NOT USE IF DAMAGED.**
- **Do not use** silicone-based lubricants on this catheter if it has a balloon.
- This catheter must only be used under medical supervision by personnel who have been adequately trained in safe gastro-intestinal intubation, and in the performance of perfused manometric measurements and/or barostat measurements.
- Post-procedure manometric catheters are contaminated with body fluid. Immediately dispose of device to prevent re-use.
- Manufacturer's Warranty – 6 months from Manufactured Date for any manufacturer's defects (i.e. defects in materials and/or workmanship).
- **MUI SCIENTIFIC IS NOT LIABLE FOR ANY DAMAGE TO THE CATHETER, OR HARM TO THE PATIENTS OR PERSONNEL, CAUSED BY IMPROPER USE.**



Edit date: February 19, 2019



## INSTRUCTION FOR USE FOR SINGLE-USE MANOMETRIC WATER-PERFUSED AND/OR BAROSTAT CATHETERS

- This catheter is designed for **Single-Use Only**. Attempts to reprocess, resterilize, and/or re-use may lead to failure of the device and/or cross-contamination.
- Inspect catheter for any evidence of mechanical damage or imperfections. **DO NOT USE IF DAMAGED.**
- **Do not use** silicone-based lubricants on this catheter if it has a balloon.
- This catheter must only be used under medical supervision by personnel who have been adequately trained in safe gastro-intestinal intubation, and in the performance of perfused manometric measurements and/or barostat measurements.
- Post-procedure manometric catheters are contaminated with body fluid. Immediately dispose of device to prevent re-use.
- Manufacturer's Warranty – 6 months from Manufactured Date for any manufacturer's defects (i.e. defects in materials and/or workmanship).
- **MUI SCIENTIFIC IS NOT LIABLE FOR ANY DAMAGE TO THE CATHETER, OR HARM TO PATIENTS OR PERSONNEL, CAUSED BY IMPROPER USE.**



Edit date: February 19, 2019